



## Centar za mikrobiologiju

**Datum:** 27.09.2017.

**Del.broj:** 995/17

**Predmet:** Pojašnjenje tenderske dokumentacije za nabavku br.2/17

### *Pitanje i Odgovor*

Postovani, molimo Vas da nam u cilju izrade sto kvalitetnije ponude a u skladu sa zahtevima Vase tenderske dokumentacije za nabavku broj 2/17, odgovorite na sledeca pitanja.

1. Da li je u partiji 4 odgovarajuće ponuditi aparat prema specifikaciji:

### Microincinerator Nova

Cost-efficient infrared sterilizer for needles and inoculating loops made out of metal.

- For sterilizing without using a flame in the open air
- Eliminates the aerosol effects of microorganisms
- Infrared heating in a ceramic tube protects users from any gases or flames
- At a temperature of 815 °C, the needle is sterilized in 5 to 7 seconds
- Heating element can be adjusted in 5 different positions
- The heating element is protected by a perforated stainless steel case



2. Da li je u partiji 3 odgovarajuće ponuditi aparat prema specifikaciji:



## LLG Min/Max Thermometer with outdoor sensor

Digital thermometer with outdoor sensor.

- Two-tier LCD display
- Measuring temperature range  $-50^{\circ}\text{C}$  to  $+70^{\circ}\text{C}$
- Display resolution  $0.1^{\circ}\text{C}$
- Accuracy  $\pm 1^{\circ}\text{C}$
- Dimensions 110 x 70 x 20mm
- Display size 40 x 30mm
- AAA 1.5V battery included
- Length of the cable 190cm



3. Da li je za partije 3 i 4 odgovarajuće dostaviti ISO 9001 ovlašćenog distributera čiju autorizaciju dostavljemo uz ponudu?

1. Потребно је за партију 4 понудити апарат који у потпуности одговара спецификацији објављеној на порталу.

2. Последња карактеристика апарата из партије 3 који Ви нудите не одговара спецификацији апарата која је објављена на порталу (дужина кабла је 190 cm а потребно је 300 cm).

3. Неопходно је, као што је наведено у спецификацији, доставити ISO 9001 произвођача траженог апарата.

Tehnički rukovodilac Centra za mikrobiologiju

Službenik za javne nabavke

*Kodrić Jadranka*

Nataša Teleki